

NUCLEAR SAFETY AND RADIATION PROTECTION ACT

(1995 No. 19)

NIGERIAN RADIATION SAFETY IN RADIOTHERAPY REGULATIONS, 2006



ARRANGEMENT OF REGULATIONS

REGULATION

PART I—GENERAL

1. Interpretation.
2. Objective.
3. Scope.

PART II—PRINCIPAL REQUIREMENTS

4. Authorization of the practice.
5. Renewal of authorization.
6. Personnel Accreditation for radiation protection and safety.
7. Authorization of other practices related to radiotherapy.
8. Revocation of authorization.
9. Withdrawal of authorization.
10. Inspection.
11. Radiation safety and protection requirements.

PART III—MANAGERIAL REQUIREMENTS

12. Managerial commitment and policy statement.
13. Organization and responsibilities.
14. Establishment of Radiation Protection Programmes.
15. Quality assurance.
16. Human factor: Staffing.
17. Education qualification and training.

PART IV—SAFETY OF SOURCES

18. Prevention and mitigation of accidents.
19. Safety in the design of radiation sources and equipment.
20. Sealed sources.
21. Equipment.
22. Design requirements.
23. Safety systems

24. Manual brachytherapy.
25. Remote control brachytherapy and external beam therapy.
26. Safety associated with acceptance test, commissioning and operation.
27. Safety associated with commissioning.
28. Safety associated with operation.
29. Maintenance.
30. Safe operation of external beam therapy.
31. Safe operation of brachytherapy.
32. Safe operation of manual brachytherapy.
33. Safe operation of remote control afterloading brachytherapy.
34. Security of sources.

PART V—OCCUPATIONAL EXPOSURE

35. Responsibilities and conditions of service.
36. The use of dose constraints in Radiotherapy.
37. Investigation levels in radiotherapy.
38. Pregnant workers.
39. Classification of areas.
40. Local rules and supervisions.
41. Protective equipment and tools.
42. Individual exposure monitoring.
43. Investigation and follow-up.
44. Monitoring of workplace.
45. Protection of workers in interventions.
46. Health surveillance.
47. Records.

PART VI—MEDICAL EXPOSURE

48. Responsibilities.
49. Justification.
50. Optimization.
51. Calibration.
52. Clinical dosimetry.
53. Quality assurance for medical exposures.
54. Discharge of patients.
55. Investigation of accidental medical exposure.

PART VII—PUBLIC EXPOSURE

56. Responsibilities.
57. Access control of visitors.
58. Radioactive waste and sources no longer in use.
59. Monitoring of public exposure.

PART VIII—POTENTIAL EXPOSURE, MITIGATION AND EMERGENCY PLANS

60. Potential exposure : Safety assessment.
61. Accident prevention.
62. Mitigation : Emergency plans.


PART IX—SAFETY IN TRANSPORTATION OF RADIOACTIVE MATERIALS

63. Safety in transportation of radioactive materials.
64. Receipt of radioactive materials.
65. Dispatch of radioactive materials.
66. Empty packages.
67. Return of disused sources.

PART X—OFFENCES, PENALTIES AND APPEALS

68. Offences and penalties.
69. Appeal.
70. Citation.

SCHEDULE**DOSE LIMITS FOR OCCUPATIONAL AND PUBLIC EXPOSURE**



NUCLEAR SAFETY AND RADIATION PROTECTION ACT
(1995 No. 19)

NIGERIAN RADIATION SAFETY IN RADIOTHERAPY REGULATIONS, 2006

[5th October, 2006]

Commence-
ment.

In exercise of the powers conferred on it by Section 47 of the Nuclear Safety and Radiation Protection Act 1995 and of all other powers enabling it in that behalf, THE NIGERIAN NUCLEAR REGULATORY AUTHORITY, with the approval of the President, hereby makes the following Regulations -

PART I—GENERAL

1. These regulations, unless the context otherwise requires :

Interpreta-
tion.

"absorbed dose" means the fundamental dosimetric quantity D, defined as :

$$D = \frac{d\varepsilon}{dm}$$

where $d\varepsilon$ "means" energy imparted by ionizing radiation to matter in a volume element and dm is the mass of matter in the volume element. The energy can be averaged over any defined volume, the average dose being equal to the total energy imparted in the volume divided by the mass in the volume. The SI unit of absorbed dose is the joule per kilogram (J/kg), termed the gray (Gy) ;

"accident" means any unintended event, including operating errors, equipment failures or other mishaps, the consequences or potential consequences of which are not negligible from the point of view of protection or safety ;

"Act" means Nuclear safety and Radiation Protection Act 1995 ;

"afterloading" means technique by which applicators or guides are placed on the patient prior to placement of the radioactive material permitting verification of correct positioning as well as prompt loading and unloading of radioactive material ;

"applicant" means any legal person who applies to the Authority for authorization to undertake any of the actions described in the NiBIRR ;

"approved" means approved by the Authority ;

"Authority" means Nigerian Nuclear Regulatory Authority ;

"authorization" means a permission granted in a document by Authority to a legal person who has submitted an application to carry out a practice or any other action described in the Act and NiBIRR. The authorization can take the form of a registration or a licence ;

"authorized" means granted an authorization by Authority ;

"controlled area" means any area in which specific protection measures and safety provisions are or could be required for :

(a) controlling normal exposures or preventing the spread of contamination during normal working conditions ; and

(b) preventing or limiting the extent of potential exposures ;

"dose constraint" means a prospective and source related restriction on the individual dose delivered by the source which serves as a bound in the optimization of protection and safety of the source. For occupational exposures, dose constraint is a source related value of individual dose used to limit the range of options considered in the process of optimization. For public exposure, the dose constraint is an upper bound on the annual doses that members of the public should receive from the planned operation of any controlled source. The exposure to which the dose constraint applies is the annual dose to any critical group, summed over all exposure pathways, arising from the predicted operation of the controlled source. The dose constraint for each source is intended to ensure that the sum of doses to the critical group from all controlled sources remains within the dose limit. For medical exposure the dose constraint levels should be interpreted as guidance levels, except when used in optimizing the protection of persons exposed for medical research purposes or of persons, other than workers, who assist in the care, support or comfort of exposed patients ;

"dose limit" means the value of the effective dose or the equivalent dose to individuals from controlled practices that shall not be exceeded ;

"effective dose" means the quantity E, defined as a summation of the tissue equivalent doses, each multiplied by the appropriate tissue-weighting factor :

$$E = \sum_T w_T \cdot H_T$$

where H_T is the equivalent dose in tissue T and w_T is the tissue weighting factor for tissue T. From the definition of equivalent dose, it follows that :

$$E = \sum_T w_T \sum_R w_R \cdot D_{T,R}$$

where w_R is the radiation weighting factor for radiation R and $D_{T,R}$ the average absorbed dose in the organ or tissue T. The unit of effective dose is $J \cdot kg^{-1}$, termed the sievert (Sv) ;

"emergency plan" means a set of procedures to be implemented in the event of an accident ;

"employer" means a legal person with recognized responsibility, commitment and duties towards a worker in his or her employment by virtue of mutually agreed relationship. (a self-employed person is regarded as being both an employer and a worker) ;

"equipment for remote afterloading" means equipment by which treatment is delivered using remote loading devices ;

"Ethical Review Committee" means a committee of independent persons to advise on the conditions of exposure and the dose constraints to be applied to the medical exposure of individuals exposed for biomedical research purposes when there is no direct benefit to the exposed individual ;

"exclusion" means any exposure whose magnitude or likelihood is essentially unamenable to control through the requirements of these regulations is deemed to be excluded from the regulations ;

"health professional" means an individual who has been accredited through appropriate national procedures to practice a profession related to health (e.g., medicine, dentistry, chiropractic, pediatric, nursing, medical physics, radiation and nuclear medical technology, radiopharmacy, occupational health) ;

"health surveillance" means medical supervision intended to ensure the initial and continued fitness of workers for their intended task ;

"high energy radiotherapy equipment" means X-ray equipment and other types of radiation generators capable of operating at generating potentials above 300 kV, and radionuclide teletherapy equipment ;

"imaging devices" means electronic equipment used for imaging in diagnostic radiology and Nuclear medicine (e.g., image convertors, gamma cameras) ;

"interstitial application" means radiotherapeutic technique in which the radioactive material is implanted in the tumour ;

"intervention" means any action intended to reduce or avert exposure or the likelihood of exposure to sources which are not part of a controlled practice or which are out of control as a consequence of an accident ;

"intracavitary application" means radiotherapeutic technique in which the radioactive material is placed within natural cavities ;

"legal person" means any organization, corporation, partnership, firm, association, trust, estate, public or private institution, group, political or administrative entity or other persons designated in accordance with national legislation, who or which has responsibility and authority for any action taken under Nigeria Basic Ionizing Radiation Regulations ;

"licence" means an authorization granted by the Authority on the basis of a safety assessment and accompanied by specific requirements and conditions to be complied with by the Licensee ;

"licensee" means the holder of a current licence granted for a practice or source who has recognized rights and duties for the practice or source, particularly in relation to protection, safety and security ;

"medical exposure" means exposure incurred by patients as part of their own medical or dental diagnosis or treatment; by persons, other than those occupationally exposed, knowingly while voluntarily helping in the support and comfort of patients; and by volunteers in a programme of biomedical research involving their exposure;

"member of the public" means in a general sense, any individual in the population except, for the purposes of the NiBIRR, when subject to occupational or medical exposure. For the purpose of verifying compliance with the annual dose limit for public exposure, the representative individual in the relevant critical group ;

"monitoring" means the measurement of dose or contamination for reasons related to the assessment or control of exposure to radiation or radioactive substances, and the interpretation of the results ;

"NiBIRR" means Nigeria Basic Ionizing Radiation Regulations ;

"*normal exposure*" means an exposure, which is expected to be received under normal operating conditions of an installation or a source, including possible minor mishaps that can be kept under control ;

"*notification*" means a document submitted to the Authority by a legal person to notify an intention to carry out a practice or any other action described by the Authority ;

"*occupational exposure*" means all exposures of workers incurred in the course of their work with the exception of exposures excluded from the NiBIRR and exposures from practices or sources exempted by the NiBIRR ;

"*planning target volume*" means a geometrical concept used in radiotherapy for planning treatment with consideration of the net effect of movements of the patient and of the tissues to be irradiated, variations in size and shape of the tissue, and variations in beam geometry such as beam size and beam direction ;

"*potential exposure*" means exposure that is not expected to be delivered with certainty but that may result from an accident at a source or owing to an event or sequence of events of a probabilistic nature, including equipment failures and operating errors ;

"*practice*" means any human activity that introduces additional sources of exposure or exposure pathways or extends exposure to additional people or modifies the network of exposure pathways from existing sources, so as to increase the exposure or the likelihood of exposure of people or the number of people exposed ;

"*preparation table*" means working plane surface located within the storage room of a design adequate for safe handling of radioactive sources ;

"*protection and safety*" means the protection of people against exposure to ionizing radiation or radioactive substances and the safety of radiation sources, including the means for achieving such protection and safety, such as the various procedures and devices for keeping people's doses and risks as low as can reasonably be achieved and below prescribed dose constraints, as well as the means for preventing accidents and for mitigating the consequences of accidents should they occur ;

"*protective action*" means an intervention intended to avoid or reduce doses to members of the public in chronic or emergency exposure situations ;

"*public exposure*" means exposure incurred by members of the public from radiation sources, excluding any occupational or medical exposure and the normal local natural background radiation but including exposure from authorized sources and practices and from intervention situations ;

"*qualified expert in Radiotherapy Physics (medical physicist)*" means an individual who, by virtue of certification by appropriate boards or societies, professional licenses or academic qualifications and experience, is duly recognized as having expertise in radiotherapy medicine physics ;

"*quality assurance*" means all those planned and systematic actions necessary to provide adequate confidence that a product or service will satisfy given requirements for quality ;

"Radiation Safety Officer" means an individual technically competent in radiation safety and protection matters relevant for a given type of practice who is designated by the registrant or licensee to oversee the implementation of NiBIRR ;

"Radiotherapy" means treatment technique in which the radioactive material is placed within or very close to the tissue volume to be treated ;

"Radiotherapy Physician (Radiotherapist)" means specialist in radiation oncology, accredited for consulting, prescribing and supervising radiation treatment, and evaluating and summarizing the results ;

"Radiotherapy technician" means technician in radiation physics, responsible for precise delivery of the treatment prescribed by the radiotherapist, who is capable of detecting changes in patient conditions that may affect treatment ;

"Registrant" means an applicant who is granted registration of a practice or source and has recognized rights and duties for such a practice or source, particularly in relation to protection and safety ;

"reference air kerma rate" means the reference air kerma rate of a source is the kerma rate to air, in air, at a reference distance of one metre, corrected for air attenuation and scattering. This quantity is expressed in mGy h^{-1} at 1 m ;

"risk" means a multi-attribute quantity expressing hazard, danger or chance of harmful or injurious consequences associated with actual or potential exposures. It relates to quantities such as the probability that specific deleterious consequences may arise and the magnitude and character of such consequences ;

"Sealed Source" means radioactive material that is (a) permanently sealed in a capsule or (b) closely bounded and in a solid form. The capsule or material of a sealed source shall be strong enough to maintain leak tightness under the conditions of use and wear for which the source was designed, also under foreseeable mishaps.

"safety assessment" means a review of the aspects of design and operation of a source which are relevant to the protection of persons or the safety of the source, including the analysis of the provisions for safety and protection established in the design and operation of the source and the analysis of risks associated with normal conditions and accident situations.

"safety culture" means the assembly of characteristics and attitudes in organizations and individuals, which establishes that, as an overriding priority, protection and safety issues receive the attention warranted by their significance ;

"Supervised Area" means any area not designated as a controlled area but for which occupational exposure conditions are kept under review even though protective measures and safety provisions are not normally needed ;

"Unsealed Source" means a source that does not meet the definition of a sealed source ;

"source" means anything that may cause radiation exposure, such as by emitting ionizing radiation or releasing radioactive substances or materials. For example, materials emitting radon are sources in the environment, a sterilization

gamma irradiation unit is a source for the practice of radiation preservation of food, an X ray unit may be a source for the practice of diagnostic radiology ;

"Standards Dosimetry Laboratory" means a laboratory designated by the Authority for the purpose of developing, maintaining or improving primary or secondary standards for radiation dosimetry ;

"storage room" means facility designated for lodging, preparation, control and sterilization of radioactive sources ;

"supplier" means any legal person to whom a registrant or licensee delegates duties, totally or partially, in relation to the design, manufacture, production or construction of a source. (An importer of a source is considered a supplier of the source.) ;

"worker" means Any person who works, whether full time, part time or temporarily, for an employer and who has recognized rights and duties in relation to occupational radiation safety and protection. (A self-employed person is regarded as having the duties of both an employer and a worker.

Objective.

2. To protect patients, workers and the general public from the risks associated with exposure to ionizing radiation in the course of Radiotherapy practice in Nigeria and to assist licencees in meeting radiation safety requirements in radiotherapy practice for the attainment of adequate radiation protection and safety of patients, worker and the public.

Scope.

3.—(1) These regulations shall apply to all established uses of ionizing radiation sources employed in the practice of radiotherapy, to the facilities where the sources are located and to the individuals involved.

(2) These regulations shall cover occupational, public, medical, potential and emergency exposure situations.

(3) These regulations do not intend to substitute basic requirements from the Nigeria Basic Ionizing Radiation Regulations (NiBIRR).

(4) Emerging new techniques utilizing radiation sources (e.g., treatment of restenosis, and stereotactic brachytherapy) are not specifically covered by these regulations, although the general principles of protection and safety discussed are applicable.

(5) These regulations do not cover the use of unsealed sources in radiotherapy, which are covered in the Nigerian Regulations for Radiation Safety Regulations in Nuclear Medicine Regulations, 2006.

PART II — PRINCIPAL REQUIREMENTS

4.—(1) Any person who intends to utilize radiation sources for radiotherapy purposes shall notify the Authority of his intention and shall apply for an authorization.

Authoriza-
tion of the
practice.

(2) Any person applying for an authorization shall provide the Authority with the information necessary to demonstrate the safety of the practice.

(3) Considering the complexity of a radiotherapy practice, the risks involved and the fact that its safety depends on human performance, demonstration of safety requires a detailed safety assessment and therefore its authorization shall take the form of multi-stage licences.

(4) The licensee shall comply with radiation safety requirements for the following stages of the radiotherapy practice :

(a) design and construction ;

(b) operation (acceptance, commissioning, clinical use, maintenance) ;

(c) modifications ; and

(d) decommissioning (partial or total) and return or disposal of sources.

(5) Modification, with possible implications for radiation safety, of the radiotherapy facilities, of the type and activity of radioactive substances and of procedures, or cessation of the practice, shall require an amendment to the licence.

(6) Applications shall be made on prescribed Authorization form prescribed by the Authority.

5. Applications for renewal of authorization shall be submitted Three (3) months before expiration and shall be granted subject to documented evidence of due compliance with the related provisions of the Act, NiBIRR and this regulation.

Renewal of
authoriza-
tion.

6.—(1) All personnel on whom protection and safety depend shall be appropriately trained and qualified so that they understand their responsibilities and perform their duties with appropriate judgment and according to defined procedures.

Personnel
Accreditation
for radiation
protection
and safety.

(2) In radiotherapy practice, the following individuals carry responsibility for protection and safety, by virtue of tasks involving decisions, operation or manipulation of sources or equipment used in radiotherapy :

(a) medical practitioners working in radiotherapy (radiotherapists, radiation oncologists) ;

(b) qualified experts in radiotherapy physics (medical physicists) and dosimetrists or physics assistants ;

(c) other health professionals operating radiotherapy equipment or handling radioactive sources (radiotherapy technologists) ;

(d) Radiation Safety Officer ;

(e) staff for maintenance of radiotherapy equipment ; and

(f) staff performing special tasks (type tests, long term stability checks, etc.).

(3) These individuals shall provide evidence of education qualification and training relevant to their duties related to the use of radiation sources.

(4) For radiotherapists, radiation oncologists, medical physicists, dosimetrists, radiotherapy technologists and radiation safety officers, typical documentary evidence indicated above, that is, qualification credentials, shall as applicable, consist of :

(a) university degree or academic qualification relevant to the profession, issued by universities, colleges of health technology ;

(b) polytechnics and colleges of technology and other accredited federal institutions ;

(c) accreditation to exercise the profession granted by the relevant competent authorities or other professional or academic bodies recognized by the Authority ;

(d) attendance at and passing of required examinations on a course on radiation protection for which the contents, the methodology and the teaching institution are accredited by the Authority or by other professional bodies recognized by the Authority. This course may be integrated in the curricula of the professional education under (a) and (b) ; and

(e) on-the-job training supervised by professionals with accreditation by the Authority or other appropriate competent authorities

(5) The documentary evidence for an individual to perform maintenance of radiotherapy equipment shall consist of :

(a) certification by the manufacturer of his or her having completed a training programme on the type of authorized equipment (the certification should indicate the type of equipment and the parts of the equipment that the engineer or technician has been trained to repair or adjust, or the scope of the maintenance he/she is enable to perform) ; and

(b) a course on radiation protection for which contents, methodology and teaching institution are approved by the Authority.

7. Consideration that, according to the Act and the Nigeria Basic Ionizing Radiation Regulations, the activities listed below also require authorization, licensees shall contract any of the following services only to enterprises authorized by the Authority :

(a) import, distribution, sale or transfer of radioactive sources ;

(b) installation, maintenance of radiotherapy equipment, including source change, and decommissioning ;

(c) disposal of radioactive sources.

8. In the event of a breach of any licence condition, the licensee shall, as appropriate :

(a) investigate the breach and its causes, circumstances and

(b) consequences ;

(c) take appropriate action to remedy the circumstances that led to the breach and to prevent a recurrence of similar breaches ;

Authoriza-
tion of other
practices
related to
Radio
therapy.

Revocation
of authoriza-
tion.

(d) communicate to the Authority, and to any relevant organizations when applicable, on the causes of the breach and on the corrective or preventive actions taken or to be taken; and

(e) take whatever other actions necessary as required by the Authority.

9.—(1) Failure to take corrective or preventive actions within a reasonable time shall be grounds for modifying, suspending or withdrawing any authorization that had been granted by the Authority.

Withdrawal
of authoriza-
tion.

(2) The Authority will suspend or revoke an authorization when licensees are in breach of the conditions of the authorization, NiBIRR or specific requirements of these regulations.

(3) In order to be able to resume operation, the licensee shall reapply for authorization in case of revocation, or apply for reconsideration in case of suspension.

10. The Licensee shall permit inspection by the Authority of the facilities and records as often as required by the Authority.

Inspection.

11. The radiation safety and protection requirements of justification of the practice, dose limitation and optimization of protection set out in Nigeria Basic Ionizing Radiation Regulations shall be applied to radiotherapy. The dose limits for occupational and public exposure are as contained in Schedule 1.

Radiation
safety and
protection
require-
ments.

PART III—MANAGERIAL REQUIREMENTS

12.—(1) A safety culture shall be fostered and maintained to encourage a questioning and learning attitude to protection and safety and to discourage complacency.

Managerial
commitment
and policy
statement.

(2) To comply with this requirement, hospital management shall be committed to an effective protection and safety policy, particularly at senior level, and by clearly demonstrable support for those persons with direct responsibility for radiation safety, ie Radiation Safety Officer.

(3) The commitment shall be expressed in a written policy statement that clearly assigns prime importance to protection and safety in radiotherapy, whilst recognizing that the prime objective of medical care is the treatment and well-being of the patients. This action shall be followed by establishing a Radiation Safety Programme.

(4) Appropriate resources shall be made available to support the commitment stated in sub-paragraph 3 above.

13.—(1) The principal parties having the main responsibilities for the application of the Nigeria Basic Ionizing Radiation Regulations and these regulations shall be licensees and employers.

Organiza-
tion and
responsi-
bilities.

(2) Other parties shall have subsidiary responsibilities for the application of the Nigeria Basic Ionizing Radiation Regulations. These parties may include, as appropriate: suppliers, workers, radiation safety officers, medical practitioners, health professionals, qualified experts, ethical review committees and any other party to whom a principal party has delegated specific responsibilities.

(3) The licensee shall assign clear responsibilities to personnel (e.g., medical practitioners, qualified experts in radiotherapy physics, radiotherapy technologists, radiation safety officers and other health professionals) so that adequate radiation protection of patients, workers, and the public is ensured.

(4) The need for qualified experts shall be determined; their responsibilities defined and suitable persons shall be appointed either on full-time or part-time basis.

14.—(1) The licensee shall establish a radiation safety programme and provide the necessary resources to comply with this programme. The programme shall relate to all phases of the practice, from design through operation to decommissioning. The programme shall reflect management responsibility for radiation protection and safety through management structure, policies, procedures and organizational arrangements that are commensurate with the risks.

(2) The licensee shall appoint a Radiation Safety Officer. The Radiation Safety Officer shall have sufficient authority and to communicate with management regarding compliance with these regulations and licence provisions.

(3) A Radiation Safety Committee shall be formed in all radiotherapy institutions that handles Category 1 and 2 sources (see Nigerian Safety and Security Regulations, 2006). The committee shall co-ordinate and review the radiation safety and protection programme and quality assurance procedures. The scope of the committee shall also cover other practices using ionizing radiation in the hospital.

15. The licensee shall establish a comprehensive Quality Assurance Programme for radiation protection and safety to ensure that all necessary procedures are developed and implemented in order to comply with the regulations for radiation safety within the terms and conditions of the authorization.

16. The licensee shall appoint a number of professionals, with personal accreditation for the tasks described in these regulations, sufficient to ensure that all activities relevant to protection and safety are carried out in accordance with these regulations, and the number of persons shall be kept under review, especially as workload increases, or new techniques and new equipment are incorporated.

17.—(1) All staff working with radiation sources or treating patients shall have an adequate educational background with relevant practical training, relevant to their duties.

(2) The licensee shall ensure that only staff with the credentials specified in these regulations fill these positions and that they are aware of :

- (a) the conditions and limitations of the licence ;
- (b) the institutional radiation safety and protection policies and
- (c) procedures (including practice recover) ;
- (d) their own individual responsibilities ;
- (e) the use and operation of equipment ;
- (f) the local quality assurance programme and quality control procedures, which shall be in an accessible manual recover) ;

Establishment of radiation protection programmes.

Quality Assurance.

Human factors: Staffing

Educational qualification and training of Staff

(g) review and analysis of incidents and accidents that have occurred in the institution or documented from elsewhere ; and

(h) instructions provided to patients and caregivers.

(3) The professional education and the training mentioned above shall be completed before commencement of duties and continued subsequently as part of the professional development and as required by the Authority. Furthermore, the instruction of personnel is required whenever significant changes occur in duties, regulations, the terms of the licence or radiation safety procedures.

(4) The licensee shall ensure that the following staff are provided with specific instructions on radiation safety and protection :

(a) brachytherapy nurses ;

(b) staff who do not belong to radiotherapy practice but need to enter controlled areas ; and

(c) staff who transport radioactive materials within the institution.

(5) The licensee shall establish a policy that encourages and provides continuing professional development programme, with the aim of improving staff skills, maintain familiarity with current practices and foster a safety culture throughout the institution.

(6) Personnel with duties in the vicinity of radioactive sources used in radiotherapy shall be informed of the radiation hazard, details of the specific uses, and the radiation safety programme.

(7) The licensee shall keep a record of the initial and periodic instruction of personnel as part of the records of the Radiation Safety Committee.

PART IV—SAFETY OF SOURCES

18. A multilayer defence-in-depth system of provisions for protection and Safety commensurate with the magnitude and likelihood of the potential exposures involved shall be applied to sources such that a failure at one layer is compensated for or corrected by subsequent layers, for the purposes of :

Prevention and mitigation of accidents

(a) preventing exposure that may cause accidents ;

(b) mitigating the consequences of any such accident that does occur ; and

(c) restoring sources to safe conditions after any such accident.

19.—(1) The requirements for the safety of sources specified in Nigeria Basic Ionizing Radiation Regulations in general shall also apply to sources used in radiotherapy, where relevant, and, in particular, equipment used in radiotherapy shall be so designed that :

Safety in the design of radiation sources and equipment.

(a) failure of a single component of the system be promptly detectable so that any unplanned medical exposure of patients is minimized ; and

(b) the incidence of human error in the delivery of unplanned medical exposure be minimized.

(2) Licensees shall :

(a) taking into account information provided by suppliers, identify possible equipment failures and human errors that could result in unplanned medical exposures ;

(b) take all reasonable measures to prevent failures and errors, including the selection of suitably qualified personnel, the establishment of adequate procedures for the calibration, quality assurance and operation of therapeutic equipment, and the provision to personnel of appropriate training and periodic retraining in the procedures, including protection and safety aspects ;

(c) take all reasonable measures to minimize the consequences of failures and errors that may occur ; and

(d) develop appropriate emergency plans for responding to events that may occur, display plans prominently and periodically conduct practice drills.

20.—(1) Radioactive sources for either teletherapy or brachytherapy shall be so constructed that they conform to the definition of a sealed source.

(2) To meet this requirement, sealed sources used for external beam therapy and brachytherapy shall comply with International Standard Organization 2919 or the Standard Organization of Nigeria equivalent standards.

(3) Sealed sources used for external beam and brachytherapy shall have a calibration certificate provided by the manufacturer, in accordance with International Standard Organization and Standard Organization of Nigeria 1677.

(4) The use of older teletherapy units containing ^{137}Cs and brachytherapy sources incorporating ^{226}Ra or old ^{137}Cs in preloaded applicators is no longer justified and shall no longer be used.

(5) Sources used for manual brachytherapy shall be used either with metallic applicators or with plastic applicators compatible with the sources as stated by the manufacturer.

(6) In brachytherapy, when applicators are employed, they shall be those manufactured specifically for the source or those, which are compatible. Licensees shall not use radioactive sources after their manufacturer-recommended working life.

(7) Sources using beta emitters shall be provided with shielding of low atomic number to minimize bremsstrahlung, while in storage and preparation for use.

21.—(1) As established by these regulations, radiation sources including radioactive material, equipment and accessories, shall be purchased only from authorized suppliers and shall have a valid type test. Written methods for the purchase, installation, acceptance, commissioning, use, maintenance and quality control of such material shall be developed with the involvement of the responsible staff or the Radiation Safety Committee.

(2) Licensees, in specific co-operation with suppliers, shall ensure that, with regard to equipment consisting of radiation generators and that containing sealed sources used for medical exposures:

(a) whether imported into or manufactured in the country where it is used, the equipment conform to applicable standards of the International Electrotechnical Commission and the International Standards Organization or to Standard Organization of Nigeria.

Sealed
sources.

Equipment.

(b) performance specifications and operating and maintenance instructions, including protection and safety instructions, shall be provided in English and in compliance with the relevant international Electrotechnical Commission or International Standards Organization standards with regard to 'accompanying documents' ; and

(c) where practicable, the operating terminology (or its abbreviations) and operating values be displayed on operating consoles in English.

(3) Compliance with International Electrotechnical Commission and Standard Organization of Nigeria standards shall be demonstrated and supported by written evidence,

(4) Compliance shall be confirmed for the particular piece of equipment delivered, by including the relevant tests of the International Electrotechnical Commission standards in the acceptance protocol. The set of tests to be included in the protocol shall be specified in the purchasing conditions.

(5) The Requirements for radiation generators and irradiation installations for radiotherapy shall be, in pertinent part :

(a) radiation beam control mechanisms shall be provided, including devices that indicate clearly and in a fail-safe manner whether the beam is 'on' or 'off' ;

(b) as nearly as practicable, the exposure shall be limited to the area being examined or treated by using collimating devices aligned with the radiation beam ; and

(c) the radiation field within the examination or treatment area without any radiation beam modifiers (such as wedges) shall be as uniform as practicable and the non-uniformity shall be stated by the supplier; and exposure rates outside the examination or treatment area due to radiation leakage or scattering be kept as low as reasonably achievable.

(6) Licensees, in specific co-operation with suppliers, shall ensure that :

(a) radiation generators and irradiation installations include provisions for selection, reliable indication and confirmation (when appropriate and to the extent feasible) of operational parameters such as type of radiation, indication of energy, beam modifiers such as filters, treatment distance, field size, beam orientation and either treatment time or preset dose ;

(b) therapy installations using radioactive sources shall be fail-safe in the sense that the source will be automatically shielded in the event of an interruption of power and will remain shielded until the beam control mechanism is reactivated from the control panel ;

(c) have at least two independent 'fail to safety' systems for terminating the irradiation ;

(d) the equipment is safety interlocks or other means designed to prevent the clinical use of the machine in conditions other than those selected at the control panel, are provided ; and

(e) the design of safety interlocks shall be such that operation of the installation during maintenance procedures, if interlocks are bypassed, could be performed

only under direct control of the maintenance personnel using appropriate devices, codes or keys.

(7) (a) Teletherapy equipment containing radioactive sources shall be provided with a device to return sources manually to the shielded position in case of emergency. For gamma knife units, it shall be possible to close the shielding door manually; and

(b) The irradiation heads in external beam equipment and source containers in brachytherapy shall be provided with a clear permanent sign indicating the existence of radioactive material (i.e., ISO 361 symbol). Containers and devices containing radioactive sources, when outside the radiotherapy department, shall be labeled with a warning signal, which is recognized as "DANGER" by any member of the public.

22.—(1) The following design requirements are applicable only to external beam therapy equipment:

(a) the primary beam shall be only directed towards primary barriers with sufficient shielding. If a primary shielding is incorporated to the equipment, electrical or mechanical interlocks shall be provided to avoid that the beam be directed toward the secondary barriers when the shielding is not intercepting the beam;

(b) external beam teletherapy equipment shall be stable in any position and it shall be possible to fix it in any desired position; and

(c) couch and table top movements (vertical, longitudinal, lateral, angular) shall facilitate patient positioning and set-up and shall be fixed by adequate brakes.

23.—(1) The design of the facility shall make provisions for safety systems or devices to be inherent to the equipment and room, to lower the probability of occurrence of abnormal situations.

(2) The shielding shall be calculated using optimization principles and, dose constraints shall be developed and used as required by these regulations. The overall design of the facility including these calculations shall be performed by an appropriate qualified expert.

(3) The licensee shall consider access control when determining the location of therapy treatment rooms and source storage areas.

(4) Radiation monitoring equipment, where appropriate, shall be installed or be available to give warning of an unusual situation in the use of radiation generators and radionuclide therapy equipment.

24.—(1) The following rooms shall be available for manual brachytherapy:

(a) source storage and preparation laboratory;

(b) an operating room; and

(c) a treatment planning room.

(2) Access to nursing, i.e., a nursing station next to the wards, shall be available as it has an impact on safety.

Design
require-
ments.

Safety
systems.

Manual
brachytherapy.

(3) The source storage and preparation laboratory shall have a link for cleansing of sources, provide a filter or trap suitable for preventing loss of sources through the sewage.

(4) The storage and preparation of radioactive sealed sources for manual brachytherapy shall meet the following requirements :

(a) the room shall only be used for source storage and preparation by designated and trained personnel ;

(b) the room shall be provided with a locked door to control access and maintain source security ;

(c) a radiation sign and danger warning sign shall be posted on the door ;

(d) there shall be shielded safe storage for all sources. The outer surface of the storage shall be made of fire-proof materials. The safe shall be located near the preparation workbench to reduce the exposure of personnel during transfer of sources;

(e) the safe shall have compartments for different source activities. Each compartment shall be marked so as to permit immediate and easy identification of its contents from the outside ;

(f) the workbench shall be provided with an L-block shielding with a lead glass viewing window ;

(g) the source handling area shall be well illuminated and a magnifying glass in a fixed mounting shall be available ;

(h) devices for handling sources, typically forceps, shall be available. They shall be as long as practicable, compatible with the effective manipulations. A device shall be provided for threading sources expeditiously with the fingers protected by distance ;

(i) sources shall be readily identifiable by sight and where radioactive sources of the same appearance but of different activities are used, they shall be distinguishable, e.g., by different coloured threads or beads ;

(j) hand carried transport containers shall be provided with long handles and the lid of the container shall be securely fastened to prevent spillage of sources during transportation and containers shall bear the radiation symbol as well as a danger warning sign ;

(k) there shall be a clear indication of the radiation level and shall be achieved either by an area radiation monitor which shall be visible on entering the room and during any handling of the unshielded sources, or a survey meter shall be available and in use during source handling ;

(l) space shall be available for source transportation trolleys with source containers ;

(m) space shall be available for secure storage to enable decay of short half-life sources, such as ^{192}Ir ; and

(n) working surface for source preparation shall be smooth and seamless to avoid losing small sources such as ^{192}Ir wires fragments.

(5) Treatment rooms shall be located in areas of low personnel flow and access shall be restricted. It is preferable that patient rooms be for single patients. If there is more than one patient room, they shall be adjacent and an area monitor shall be placed at the entrance so as to detect where a source or patient with a source is leaving the room or controlled area.

(6) Shielding shall be provided for nurses and visitors of brachytherapy patients, for which movable shields may be used within patient rooms, especially in manual brachytherapy.

(7) Prior to each treatment, movable shields shall be placed close to the patient's bed in a way that exposure to the nurses caring for the patient is minimized.

(8) The treatment room shall contain a shielded container and a remote handling tool in the event of a dislodged source.

(9) In order to ensure that no source remains within the patient, clothes or bed linen or in the area after treatment shall be monitored.

(10) External beam therapy and high dose rate brachytherapy shall be carried out in a treatment room designated for that purpose.

Remote
control
brachytherapy
and external
beam
therapy.

25.—(1) Entry to the irradiation room while the source is exposed shall be avoided with an interlock system capable of stopping the irradiation when the door is opened or unlocked.

(2) Activation of the interlock system shall automatically interrupt the treatment and the mechanism and procedures shall be capable of maintaining irradiation interruption until the door is closed and locked, and verification has been made that no person but the patient is inside the room.

(3) After an interruption, provided no operating parameters are changed or reselected, it shall be possible to restart irradiation, but only from the equipment control panel.

(4) An installed radiation monitor and/or a portable survey instrument shall be used to confirm the safe condition of the source.

(5) One or more 'off' switches shall be conveniently placed inside the treatment room to allow interruption of the irradiation from inside the room.

(6) Before start of operation, licensees shall ensure proper operation of door interlocks and other physical parameters necessary for the protection of workers and the public and for delivery of the dose to the patient.

(7) Access to the irradiation room shall be furnished with a conspicuous visible signal indicating radiation source 'on' or 'off' and the same signals for radionuclide therapy machines shall be located on the treatment head or in close vicinity.

(8) Gamma-ray therapy treatment rooms shall be provided with a stationary radiation monitor, visible on entering the room and close to a control panel with warning signal and the monitor shall be power-fail-safe in case of power interruptions or failures.

(9) The control panel shall be installed in such a way that the operator will have total overview of the access to the irradiation room at all times.

(10) Adequate systems, devices, or other means shall be provided to allow the operator to have a clear and full view of the patient and the systems for patient observation shall be redundant and independent.

(11) Oral communication shall be possible within the treatment rooms and patients, by using an intercom or other communication system.

(12) Fire fighting equipment shall be available in order to preserve source integrity in the event of fire.

26.—(1) Acceptance tests and commissioning shall not be restricted to radiation emitting equipment or sources but shall also be conducted for any system that has implications on safety, such as treatment planning systems.

Safety associated with acceptance test.

(2) After equipment installation, acceptance testing shall be conducted in order to verify that the equipment conforms to technical specifications given by the manufacturer and to verify compliance with safety requirements of International Electrotechnical Commission Standard.

(3) As indicated in Part (2), the tests to be included in the acceptance protocol shall be specified in the purchasing conditions and contracts shall clearly establish responsibility of suppliers for resolving non-conformity identified during acceptance testing.

27. After acceptance and before starting operation, calibration of radiation sources and radiation beams and commissioning shall be performed and during commissioning, the qualified expert in radiotherapy physics shall measure all data required for clinical use.

Safety associated with commissioning.

28.—(1) Equipment shall be operated in accordance with the technical documents, ensuring satisfactory operation at all times, in respect of both, the tasks to be accomplished and radiation safety and in particular, the manufacturer's operating manual, and any additional procedures shall be approved in accordance with the quality assurance system.

Safety associated with operation.

(2) Sealed sources shall be subject to leak tests, prior to the first use and at regular intervals thereafter, in conformity with International Standard Organisation (ISO) and Leak tests shall be capable of detecting the presence of 0.2 kBq of removable contamination of the sealed source.

(3) For manual brachytherapy sources the typical method is the direct wet wipe test, while for external beam therapy and remote control brachytherapy the method to be used is the indirect wipe test of the nearest accessible surface and the sterilization process in brachytherapy shall be appropriate for preventing damage of sources and applicators that could affect safety.

(4) Quality controls need to be carried out following formally established quality control protocols, periodically, after the source has been installed or replaced, or after repairs or maintenance work that has the potential to alter the radiation output.

(5) An independent audit of the calibration of the sources shall be carried out before starting clinical use of the source, e.g., the International Atomic Energy Agency, World Health Organisation postal quality audit for dosimetry and the requirements on quality assurance for medical exposure are also provided in these regulations.

Maintenance.

29.—(1) The licensee shall ensure that adequate maintenance and inspection is performed as necessary to ensure that radiation sources retain their design requirements for radiation protection and safety for their useful lives and this requires that the licensee establish the necessary arrangements and co-ordination with the manufacturer's representative before initial operation.

(2) The licensee shall ensure that removal from and return to clinical service of radiotherapy equipment for maintenance or source exchange :

(a) is documented and a record kept ; and

(b) where maintenance of the therapy equipment or treatment planning equipment may affect the accuracy of the physical or clinical dosimetry or the safe operation of the equipment, a qualified expert in radiotherapy physics shall assess whether any specific tests or measurements are to be made and whether that equipment is operating satisfactorily before it is used to treat patients.

(3) If the licensee needs to implement an emergency plan where equipment is out of service for maintenance or source change, usually performed by another company, the licensee shall provide the radiotherapy staff with written procedures containing the scope and a description of responsibilities for these actions.

30. Safe operations of external beam treatment units require procedures for area surveys, interlock checks, wipe tests and procedures for emergencies such as a source which becomes stuck in or partially on position and the necessary equipment for procedures mentioned above shall be available, calibrated and in working order. These include :

(a) a radiation monitor with scales from mSv and greater ;

(b) wipe test capabilities ; and

(c) personal alarm dosimeters, especially for emergency intervention.

31.—(1) The source strength shall be determined individually, before it is used on a patient and the source documentation shall be checked carefully.

(2) The unit of activity used for source calibration shall be the same as the unit of activity used in the treatment planning system.

(3) Low Dose Rate and High Dose Rate sources have certain operating procedures for their safe use in common :

(a) Source inventories shall be maintained which show the location and current activity of each source at the facility as well as its unique identifier and this may be either a color-coded or letter/number identifier ;

(b) Sources shall never be left on preparation surfaces and they shall be in storage, in transit, or in use.

(c) leak tests shall be performed and documented on a periodic basis ; the tests should have sufficient sensitivity to detect the presence of 0.2 kBq of removable contamination of the sealed source :

Safe operation of external beam therapy.

Safe operation of brachytherapy.

(d) for the High Dose Rate unit, the wipe tests shall only be performed on the afterloading drive assembly and transport containers since the source itself has too high a dose rate to allow this sort of test.

(e) area surveys shall be performed periodically around the source storage facilities for Low Dose Rate and High Dose Rate sources :

(f) the storage facilities shall be marked to indicate that they contain radioactive materials and how to contact the responsible radiation safety individual in the event of an emergency.

(g) the storage facilities shall be kept locked at all times ; and

(h) after every brachytherapy treatment, the patient shall be monitored with a radiation survey meter so as to ensure that no activity remains in the patient unintentionally.

(4) The following information shall be posted in the case of Low Dose Rate brachytherapy, both manual as well as remote controlled and identification of the patient, sources, date and time of insertion and removal, nursing required, time allowance for nurses and visitors, and concise instructions for unplanned source and applicator removal and for emergency and a patient with removable source in or upon his body shall not leave the room unless accompanied by a hospital attendant.

(5) The licensee shall ensure that all brachytherapy sources are removed from the patient, except in the case of permanent implants and the patient shall be monitored with a portable detector to ensure that no source remains in or on the patient.

(6) Linen, dressings, clothing, and equipment shall be kept within the room where the removal of sources takes place until all sources are accounted for and be monitored with a radiation detector as well as rubbish bins, soiled dressing bins and laundry baskets.

(7) Mobile containers and portable equipment containing radioactive sources shall be moved to a store or to a secure place when not in use.

32.—(1) After verification of the source strength, the source or source holder shall be marked with unique identifiers (for example, a pre-established colour), to facilitate visual recognition and to prevent the possibility of confusion between different sources.

(2) Containers utilized for transport of radioactive sources shall conform with the requirements established in the Nigeria Radioactive Sources Regulations and International Atomic Energy Agency's Regulations for the Safe Transport of Radioactive Material, TS - R - 1

(3) The movements of the sources from the time they leave the safe until their return shall be documented and signed by the person responsible for the move.

(4) A person shall be assigned to be in charge of accountability for the sources. The person shall keep record of source order, issuance from and return to the safe with signatures.

Safe
operation of
manual
brachytherapy.

(5) The sources shall be inspected visually for possible damage after each use by means of magnifying viewers and a leaded viewing window in a shielded work area.

(6) There shall be a diagram at the source storage safe, which has to show exact locations of each source within the safe, aids in reducing the time it takes to locate and identify a source.

(7) Sources shall only be handled with long forceps or tongs, never directly with the fingers.

(8) When transporting the sources a mobile, shielded container shall be available and the shortest route possible should be used.

(9) Sources, which come into direct contact with body tissues, require cleaning and possible sterilization after each use, which can subject the sources to possible damage from heat, abrasion, chemical attack, and mechanical stresses and these sources shall be inspected after every use.

(10) The work surfaces shall be easily cleaned and brightly lighted to make it easy to find any sources that are dropped.

(11) As indicated in these regulations, a filter or a trap shall be used in the drain to prevent loss of sources to the sewage during cleaning.

(12) Specific precaution to be observed during the cutting and handling of ^{192}Ir wires shall include ensuring that :

(a) appropriate tools and equipment, such as forceps, cutting devices, magnifying glasses and good illumination of the work surface are available and used; if ^{192}Ir wires are cut-off for immediate use, a container to hold cut lengths shall be provided and labeled ;

(b) radioactive waste is collected and stored in adequate containers; and

(c) surfaces and tools are properly decontaminated.

33.—(1) The quality control the afterloader shall include tests to be performed at the beginning of each treatment day.

(2) The couplings and transfer tubes shall be checked for High Dose Rate and it has to be done before each treatment to ensure that there is nothing to prevent the source from moving as required.

34.—(1) Sources shall be kept secured so as to prevent theft or damage and to prevent any unauthorized use by ensuring that :

(a) control of a source shall not be relinquished without compliance with all relevant requirements specified in the licence and without immediate communication to the Authority, of information regarding any decontrolled, lost, stolen or missing source ;

(b) a source shall not be transferred unless the receiver possesses a valid authorization ; and

(c) a periodic inventory of movable sources shall be conducted at appropriate intervals to confirm that they are in their assigned locations and are secured.

Safe
operation of
remote
control after
loading
brachytherapy.

Security of
Sources.

(2) Specific provisions are required for avoiding loss of control in the following situations :

- (a) storage of sources before installation ;
- (b) temporary or permanent cessation of use ;
- (c) Storage after decommissioning whilst awaiting decision on source return or disposal ; and
- (d) brachytherapy sources remaining in the patient, clothes, bed linen or treatment area.

(3) To comply with this requirement, the licensee shall develop procedures to ensure the safe transfer and movement of radioactive sources within the institution, and establish controls to prevent theft, loss, unauthorized withdrawal or damage of sources, or entrance of unauthorized personnel to the controlled areas.

(4) The licensee shall also ensure that the number of sources in a container when removing and when returning the sources is checked and that a physical inventory of all sealed sources is performed to confirm that they are present and secure in their assigned locations.

(5) The licensee shall maintain a source movement log with a record indicating the date of removal, the name of the patient and the return of the source.

(6) Radiotherapy equipment shall be provided with safety systems capable of preventing their use by unauthorized personnel and a key shall be required to energize the system, access to which shall be restricted to authorized staff only.

(7) Any loss of source shall be reported immediately to the Radiation Safety Officer who shall report it to the Radiation Safety Committee and to the Nigerian Nuclear Regulatory Authority.

(8) All linen, dressing, clothing, equipment, and trash container shall be kept within the brachytherapy patient's room, until checks have been performed and documented that sources are not attached to them.

PART V—OCCUPATIONAL EXPOSURE

35.—(1) Licensees and employers of workers who are engaged in activities involving normal exposures or potential exposure shall be responsible for :

- (a) the protection of workers from occupational exposure ; and
- (b) compliance with any other relevant requirements of the Nigeria Basic Ionizing Radiation Regulations.

(2) Employers who are also licensees shall have the responsibilities of both employers and licensees.

(3) Workers shall :

(a) follow any applicable rules and procedures for protection and safety specified by the employer or licensee ;

(b) use properly the monitoring devices and the protective equipment and clothing provided ;

Responsibilities and conditions of service.

(c) co-operate with the employer or licensee with respect to protection and safety and the operation of radiological health surveillance and dose assessment programmes ;

(d) provide for the employer or licensee such information on their past and current work as is relevant to ensure effective and comprehensive protection and safety for themselves and others ;

(e) abstain from any wilful action that could put themselves or others in situations that contravene the requirements of the Nigeria Basic Ionizing Radiation Protection ; and

(f) accept such information, instruction and training concerning protection and safety as will enable them to conduct their work in accordance with the requirements of the Nigeria Basic Ionizing Radiation Regulations and workers can by their own actions contribute to the protection and safety of themselves and others at work.

(4) If for any reason a worker is able to identify circumstances that could adversely affect compliance with the Nigeria Basic Ionizing Radiation Regulations, the workers shall as soon as practicable report such circumstances to the employer or licensee.

(5) The management shall record any report received from a worker that identifies circumstances which could affect compliance with the Nigeria Basic Ionizing Radiation Regulations, and shall take appropriate action.

(6) If workers are engaged in work that involves or could involve a source that is not under the control of their employer, the licensee responsible for the source and the employer shall co-operate by the exchange of information and otherwise as necessary to facilitate proper protective measures and safety provisions.

36. Dose constraints shall be used for optimizing protection in the planning stage for each radiation source and anticipated individual doses shall be compared with the appropriate dose constraints and only protective measures that predict doses below dose constraints shall be chosen.

37.—(1) Employers and licensees shall, in consultation with workers or through their representatives, include in the local rules and procedures the values of any relevant investigation level or authorized level, and the procedure to be followed in the event that any such value is exceeded.

(2) The licensee shall conduct formal investigations, as required by the Nigerian Nuclear Regulatory Authority :

(a) an individual effective dose exceeds investigation levels ;

The use of dose constraints in radiotherapy.

Investigation levels in radiotherapy.

(b) any of the operational parameters related to protection or safety are out of the normal range established for operational conditions ;

(c) any equipment failure, severe accident or error occurs that causes, or has the potential to cause, a dose in excess of regulatory limits; and

(d) any other event or unusual circumstance occurs that causes, or has the potential to cause, a dose in excess of the regulatory limits or the operational restrictions imposed on the installation e.g. the significant change in workload or operating conditions of radiotherapy equipment.

(3) The investigation shall be initiated as soon as possible following the event, and a report written concerning its cause, including determination or verification of any doses received, corrective actions, and instructions or recommendations to avoid recurrence.

(4) The report shall be submitted to Authority and other concerned bodies as required, as soon as possible after the investigation, or as otherwise specified and kept for a specified period.

(5) In radiotherapy, a suitable quantity shall be used as investigation level e.g. the monthly effective dose or the dose to the hands for staff in manual brachytherapy.

38.—(1) A female worker shall, on becoming aware that she is pregnant, notify the employer in order that her working conditions may be modified where necessary.

Pregnant
workers.

(2) The notification of pregnancy shall not be considered a reason to exclude a female worker from work; however, the employer of a female worker who has notified pregnancy shall adapt the working conditions in respect of occupational exposure so as to ensure that the embryo or foetus is afforded the same broad level of protection as required for members of the public.

39.—(1) Areas in a radiotherapy practice shall be classified as controlled or supervised or public.

Classification
of areas.

(2) In radiotherapy practice, areas requiring specific protection measures shall be designated "controlled areas" and these shall include, at least, all irradiation rooms for external beam therapy and remote afterloading brachytherapy, operating rooms during brachytherapy procedures using real sources, brachytherapy patient rooms, radioactive source storage and handling areas.

(3) Controlled areas shall be demarcated by physical boundaries like walls or other physical barriers marked and identified with 'radiation area' signs.

(4) Areas surrounding brachytherapy patients rooms or around radioactive source storage and handling areas shall be designated "supervised" areas.

(5) All areas not designated as controlled nor supervised areas shall be such that persons inside are afforded the same level of protection as members of the public.

Local rules
and
supervision

40.—(1) Employers and licensees shall, in consultation with workers, through their representatives :

(a) establish in writing such local rules and procedures as are necessary to ensure adequate levels of protection and safety for workers and other persons ;

(b) include in the local rules and procedures the values of any relevant investigation level or authorized level, and the procedure to be followed in the event that any such value is exceeded ;

(c) make the local rules and procedures and the protective measures and safety provisions known to those workers to whom they apply and to other persons who may be affected by them ;

(d) ensure that any work involving occupational exposure are adequately supervised and take all reasonable steps to ensure that the rules, procedures, protective measures and safety provisions are observed ; and

(e) where the Authority required, designate a radiation safety officer.

(2) Employers, in co-operation with licensees, shall :

(a) provide to all workers adequate information on the health risks due to their occupational exposure, whether normal exposure or potential exposure, adequate instruction and training on protection and safety; and adequate information on the significance for protection and safety of their actions ;

(b) provide to female workers who are liable to enter controlled areas or supervised areas appropriate information on :

(i) the risk to the embryo or foetus due to exposure of a pregnant woman ; and

(ii) the importance for a female worker of notifying her employer as soon as she suspects that she is pregnant ;

(c) provide to those workers who could be affected by an emergency plan appropriate information, instruction and training ; and

(d) keep records of the training provided to individual workers.

41. Licensees shall ensure that workers are provided with suitable and adequate personal protective equipment, which meets any relevant regulations or standards and details of suitable equipment such as shielded L-block, lead glass, handling devices.

42.—(1) The employer of any worker, as well as self-employed individuals, and licensees shall be responsible for arranging for the assessment of the occupational exposure of workers, on the basis of individual monitoring where appropriate, and shall ensure that adequate arrangements are made with appropriate dosimetry service provider accredited by the Authority under an adequate quality assurance programme.

(2) For any worker who is normally employed in a controlled area, or who occasionally works in a controlled area and may receive significant occupational exposure, individual monitoring shall be undertaken where appropriate, adequate and practicable. In cases where individual monitoring is inappropriate, inadequate or not feasible, the occupational exposure of the worker shall be assessed on the basis of the results of monitoring of the workplace and on information on the locations and durations of exposure of the worker.

Protective
equipment
monitoring.

Individual
exposure
monitoring.

(3) These workers shall include radiation oncologists, qualified experts in radiotherapy physics, Radiation Safety Officer, radiotherapy technologists, source handlers, maintenance staff and any nursing or other staff who must spend time with patients who contain sources.

(4) For any worker who is regularly employed in a supervised area or who enters a controlled area only occasionally, individual monitoring shall not be required but the occupational exposure of the worker shall be assessed and this assessment shall be on the basis of the results of monitoring of the workplace or individual monitoring.

(5) The exchange of dosimeters in a radiotherapy department and receipt of the dose reports shall not exceed a period of one month.

(6) If an individual's dosimeter is lost, the licensee shall perform and document an evaluation of the dose the individual received and add it to the worker's dose record.

(7) Individual monitoring devices shall be calibrated and this calibration shall be traceable to a standards dosimetry laboratory.

43.—(1) The licensee shall conduct formal investigations, as required by the Authority, whenever :

- (a) the individual annual effective dose exceeds investigation levels :
- (b) any of the operational parameters subject to periodic quality control are out of the normal range established for operational conditions :
- (c) any equipment failure, or error takes place, which causes, or has the potential to cause an accident ; and
- (d) any other change or unusual circumstance that may cause an increase in dose, exceeding dose limits or the operational restrictions imposed on the installation.

(2) The investigation shall be initiated as soon as possible following the event, and a written report shall be —

(a) prepared concerning its cause, including determination or verification of any doses received, corrective or mitigating actions, and instructions or recommendations to avoid recurrence ; and

(b) submitted to the Authority and other concerned bodies as soon as possible.

44.—(1) Licensees shall develop programmes for monitoring the workplace.

(2) The nature and frequency of monitoring of workplaces shall be sufficient:

(a) be sufficient to enable :

- (i) evaluation of the radiological conditions in all workplaces,
- (ii) exposure assessment in controlled areas and supervised areas,
- (iii) review of the classification of controlled and supervised areas ; and

Investigation
and follow-
up.

Monitoring
of workplace

(b) depend on the levels of ambient dose equivalent and activity concentration, including their expected fluctuations and the likelihood and magnitude of potential exposures.

(3) The programmes for monitoring of the workplace shall specify :

(a) the quantities to be measured ;

(b) when and where the measurements are to be made and at what frequency ;

(c) the most appropriate measurement methods and procedures ; and

(d) reference levels and the actions to be taken if they are exceeded.

(4) Initial monitoring shall be conducted immediately after the installation of new radiotherapy equipment and after the replacement of teletherapy sources and remote-controlled brachytherapy sources and shall initial monitoring shall include measurements of radiation leakage from equipment within acceptance tests and area monitoring of occupable space around irradiation rooms.

(5) Initial monitoring of packages containing radioactive sources, upon receipt by the licensee, shall be performed and shall include both surveys for removable contamination and external radiation levels.

(6) Periodic monitoring shall be conducted in association with brachytherapy procedures and as soon as practicable after implantation of the sources, a survey shall be made of exposure rates in the vicinity of the patient and the transport container shall be surveyed before and after brachytherapy procedures.

(7) After removal of brachytherapy sources from a patient, a survey shall be performed to confirm removal from the patient and return to shielding of all sources and periodic monitoring of the source storage and handling area shall be conducted with a survey meter immediately following the removal from or return to storage of brachytherapy sources.

(8) Continuous monitoring of exposure levels shall be conducted through the use of area monitors in teletherapy and high dose-rate treatment rooms and continuous monitoring with an area monitor shall be conducted in source storage and handling areas.

(9) All survey meters used for workplace monitoring shall be calibrated and this calibration shall be traceable to a standards dosimetry laboratory.

Protection
of workers in
intervention.

45. Although dose limits for practices do not apply to interventions, the exposure of workers in interventions shall not be considered an unexpected exposure but rather it is deliberate and controlled and the dose limits for workers shall apply unless there is an overriding reason not to apply them; such as the need to save life after an accident or to prevent catastrophic conditions.

Health
surveillance.

46.—(1) The licensee shall make arrangements to provide medical surveillance for workers as specified by the Authority to assess the initial and continuing fitness of employees for their intended tasks and health surveillance programmes shall be based on the general principles of occupational health.

(2) Counselling shall be available to workers such as women who are or may be pregnant, individual workers who have or may have been exposed substantially in excess of dose limits, and workers who may be worried about their radiation exposure.

47.—(1) Employers and licensees shall maintain and preserve exposure records for each worker and the exposure records shall include information on the general nature of the work involving occupational exposure; information on doses, and the data upon which the dose assessments have been based.

(2) Where a worker is or has been occupationally exposed while in the employment of more than one employer, information on the dates of employment with each employer and the doses, exposures and intakes, in each such employment; and records of any doses due to emergency interventions or accidents, shall be distinguished from doses, during normal work.

(3) Employers and licensees shall provide for access by workers to information in their own exposure records; and give due care and attention to the maintenance of appropriate confidentiality of records.

PARJ VI — MEDICAL EXPOSURE

48.—(1) The licenses shall with regard to responsibilities for medical exposure ensure that :

Responsibilities.

(a) no patient is administered a therapeutic medical exposure unless the exposure is prescribed by a medical practitioner ;

(b) medical practitioners shall be assigned the primary task and obligation of ensuring overall patient protection and safety in the prescription of, and during the delivery of, medical exposure ;

(c) medical and paramedical personnel shall be available as needed and shall either be health professionals or have appropriate training adequately to discharge assigned tasks in the conduct of the therapeutic procedure that the medical practitioner prescribes ; and

(d) for therapeutic uses of radiation the calibration, dosimetry and quality assurance, the requirements of Nigeria Basic Ionizing Radiation Regulation shall be conducted by or under the supervision of a qualified expert in radiotherapy physics.

(2) All persons involved in delivery of medical exposure shall:

(a) follow the applicable rules and procedures for the protection and safety of patients, as specified by the licensee ; and

(b) be aware that prescription of treatment and treatment plan need to be signed by the medical practitioner prior to initiation of treatment.

(3) The licensee shall ensure that :

(a) the exposure of individuals incurred knowingly while voluntarily helping in the care, support or comfort of patients undergoing medical diagnosis or treatment are constrained as specified second schedule of Nigeria Basic Ionizing Radiation Regulations ; and

(b) the training criteria shall be specified or be subject to approval, as appropriate, by the Authority in consultation with relevant professional bodies.

(4) Medical practitioners shall promptly inform the licensee of any deficiencies or needs regarding compliance with the Nigeria Basic Ionizing Radiation Regulations and these regulations in respect of protection and safety of patients and shall take such actions as may be appropriate to ensure the protection and safety of patients.

Justification.

49.—(1) Medical exposures shall be justified by weighting the therapeutic benefits they produce against the radiation detriment they might cause, taking into account the benefits and risks of available alternative techniques that do not involve medical exposure.

(2) The licensee shall ensure that medical practitioners follow a justification procedure that is documented and signed and the medical practitioner shall consider the efficacy, benefits and risks of alternative treatment modalities, e.g., surgery and chemotherapy, either alone or in combination with radiation therapy.

(3) In respect of medical research, the exposure of humans for medical research is deemed not justified unless it is :

(a) in accordance with the provisions of the Helsinki Declaration and follow the guidelines for its application prepared by Council for International Organizations of Medical Sciences and World Health Organisation ; and

(b) subject to the advice of an Ethical Review Committee or any other institutional body assigned similar functions by the Federal Ministry of Health and to applicable national and local regulations.

(4) Exposure to comforters shall be in accordance with the dose constraint formalized in the second scheduled of Nigeria Basic Ionizing Radiation Regulations and they shall be provided with instruction on actions to take to limit their exposure while visiting or caring for a patient who has received a brachytherapy implant.

Optimization.

50.—(1) Licensees of radiotherapy practices shall ensure that :

(a) exposure of normal tissue during radiotherapy are kept as low as reasonably achievable consistent with delivering the required dose to the planning target volume, and organ shielding shall be used where practicable and appropriate;

(b) radiotherapy procedures causing exposure of the abdomen or pelvis of women who are pregnant or likely to be pregnant shall be avoided unless there are strong clinical indications ;

(c) any therapeutic procedure for pregnant women shall be planned to deliver the minimum dose to any embryo or foetus ; and

(d) the patient shall be informed of possible risks.

(2) The licensee shall provide written instructions on actions to take to reduce exposure to comforters, caregivers and members of the public from sources in brachytherapy patients with permanent implants and these instructions shall include minimizing prolonged contact with children and potentially pregnant women, and procedures to follow in the event that a source becomes dislodged.

Calibration.

51.—(1) Licensees shall ensure that :

(a) the calibration of sources used for medical exposure are traceable to Standards dosimetry laboratory ;

(b) radiotherapy equipment shall be calibrated in terms of radiation quality or energy and either absorbed dose or absorbed dose rate at a predefined distance under specified conditions, e.g. following the recommendations given in relevant International Atomic Energy Agency Technical Reports Series ;

(c) sealed sources used for brachytherapy shall be calibrated in terms of activity, reference air kerma rate in air or absorbed dose rate in a specified medium, at a specified distance, for a specified reference date ; and

(d) the calibrations shall be carried out at the time of commissioning a unit, after any maintenance procedure that may have an effect on the dosimetry and at intervals approved by the Authority.

(2) The licensee shall ensure that all teletherapy equipment outputs are compared at least once every two years in a national, regional or international programme for independent dose verification, such as the Thermo Luminescence Dosimetry (TLD) postal service established by the International Atomic Energy Agency/ World Health Organisation.

(3) The licensee shall develop or adopt, implement, and follow a protocol for calibration of radiation sources used for radiotherapy.

(4) Calibration of new equipment and new radiation sources shall be done independently by at least two different qualified experts in radiotherapy physics and preferably using different dosimetry systems and the results shall be compared only after the completion of both measurements.

(5) New brachytherapy sources shall be calibrated and an investigation shall be conducted where the measurement varies by more than 5% from the manufacturer's certified activity or kerma rate and the sources shall not be used for patient's treatment if the difference is greater than 10% until the difference is investigated and resolved.

(6) The responsibility for the investigation and for further action remains with the licensee, and the investigation is usually performed by the qualified expert in radiotherapy physics, with or without external help.

52.—Licensees shall ensure that the following items are determined and documented :

Clinical
dosimetry.

(a) for each patient treated with external beam radiotherapy equipment, the maximum and minimum absorbed doses to the planning target volume together with the absorbed dose to a relevant point such as the centre of the planning target volume, plus the dose to other relevant points selected by the medical practitioner prescribing the treatment ;

(b) in brachytherapeutic treatments performed with sealed sources, the absorbed doses at selected relevant points in each patient ; and

(c) in all radiotherapeutic treatment, the absorbed doses to relevant organs.

53.—(1) Licensees, in addition to applying the relevant requirements for quality assurance specified in the Nigeria Basic Ionizing Radiation Regulations, shall establish a comprehensive quality assurance programme for medical exposures with the participation of appropriate qualified experts in the relevant fields, such as radiophysics

Quality
assurance for
medical
exposures.

or radiopharmacy, taking into account the principles established by the World Health Organisation and the Pan American Health Organisation.

(2) Quality assurance programmes for medical exposures shall include :

(a) measurements of the physical parameters of the radiation generators, imaging devices and irradiation installations at the time of commissioning and periodically thereafter ;

(b) verification of the appropriate physical and clinical factors used in patient diagnosis or treatment ;

(c) written records of relevant procedures and results ;

(d) verification of the appropriate calibration and conditions of operation of dosimetry and monitoring equipment ; and

(e) as far as possible, regular and independent quality audit reviews of the quality assurance programme for radiotherapy procedures.

(3) Following the acceptance of new radiotherapy equipment, sufficient data shall be measured at the commissioning, to be used for clinical dosimetry in treatment planning and this data shall be clearly documented in the work book and before being issued for use in treatment planning, the documentation shall be independently verified, signed and dated.

(4) All dosimetry calibrations, clinical dosimetry data and methods of calculation for therapy equipment shall be reconfirmed at regular intervals and the measurements and checks carried out for this purpose shall be sufficiently comprehensive to detect any significant variations from the data in use.

(5) For imaging modalities used for anatomical information and for quantitative purposes in treatment planning computer systems, measurements shall be made by the qualified expert in radiotherapy physics or to determine the required physical data and the size of errors due to instrument distortions and corrections shall be made for significant errors.

(6) Verification of the treatment through "in vivo" dosimetry shall be performed and total dose to the tumor shall be estimated and compared with prescribed dose.

54.—(1) In order to restrict the exposure of any members of the household of a patient who has undergone a therapeutic procedure with sealed or unsealed radionuclides and members of the public, such a patient shall not be discharged from hospital before the activity of radioactive substances in the body falls below the level specified in the third schedule of the Nigeria Basic Ionizing Radiation Regulations.

(2) Written instructions to the patient concerning contact with other persons and relevant precautions for radiation protection and safety shall be provided where necessary.

55.—(1) Licensees shall promptly investigate any of the following incidents :

(a) any therapeutic treatment delivered to either the wrong patient or the wrong tissue, or using the wrong pharmaceutical, or with a dose or dose fractionation differing substantially from the values prescribed by the medical practitioner or which may lead to undue acute secondary effects ; and

Discharge of patients.

Investigation of accidental medical exposure.

(b) any equipment failure, accident, error, mishap or other unusual occurrence with the potential for causing a patient exposure significantly different from that intended.

(2) Licensees shall, with respect to any investigation required by these regulations :

(a) calculate or estimate the doses received and their distribution within the patient ;

(b) indicate the corrective measures required to prevent recurrence of such an incident ;

(c) implement all the corrective measures that are under their own responsibility ;

(d) submit to the Authority, as soon as possible after the investigation or as otherwise specified by the Authority, a written report which states the cause of the incident and includes the information specified in (regulation 55 (2) paragraph a-c, as relevant, and any other information required by the Authority ; and

(e) inform the patient and his or her doctor about the incident.

PART VII — PUBLIC EXPOSURE

56.—(1) The licensee is responsible for controlling public exposure resulting from a radiotherapy practice.

Responsibilities

(2) Access by members of the public to areas in and near the radiotherapy department shall be considered when designing shielding of storage and use of facilities.

(3) The licensee shall :

(a) develop and implement use, transport and storage measures for ensuring the safety and security of radiotherapy sources to control public exposures in accordance with the requirements of the Authority ; and

(b) control and maintain constant surveillance of sources that is not in storage and secure stored sources from unauthorized access, removal, or use, and the storage facility shall be locked at all times.

57. The licensee shall make arrangements to control access of members of the public to radiotherapy irradiation rooms, and provide adequate information and instruction to the public before they enter a controlled area so as to ensure appropriate protection (of members of the public and shall be accompanied).

Access control for visitors.

58.—(1) The licensee shall notify the Authority and submit a plan for transfer or disposal of sources if they are no longer in use.

Radioactive waste and sources no longer in use.

(2) The licensee shall be responsible for the sources until the time of their transfer to another appropriate licensee or to an authorized waste disposal facility and in particular, the licensee shall :

(a) notify the Authority of any intention to transfer or decommission ^{60}Co teletherapy equipment prior to initiating any action and depleted uranium used as shielding material shall also be treated as radioactive waste. e.g, a ^{60}Co teletherapy head may contain depleted uranium and shall be disposed of appropriately ; and

(b) ensure that resources for the disposal of the sources are be made available where the teletherapy equipment is to be decommissioned.

Monitoring of public exposure.

59.—(1) The licensee shall, as appropriate :

(a) establish and carry out a monitoring programme sufficient to ensure that the requirements of the Nigeria Basic Ionizing Radiation Regulations regarding public exposure to sources of external irradiation be satisfied and to assess such exposure ; and

(b) keep appropriate records of the results of the monitoring programmes.

(2) The programme for monitoring public exposure from radiotherapy shall include dose assessment in and the surrounding of irradiation rooms for external beam therapy, of brachytherapy wards and source storage, preparation room and waiting rooms.

PART VIII — POTENTIAL EXPOSURE, MITIGATION AND EMERGENCY PLANS

Potential exposure : Safety assessment.

60.—(1) Licensees shall prepare a safety assessment applied to all stages of design, construction, operation, maintenance and decommissioning of the radiotherapy facility, and submit it to the Authority.

(2) The assessment shall be systematic and contain information on identification of possible events leading to accidental exposure.

(3) The safety assessment shall not only cover these events, but also aim at anticipating other events that have not previously been reported.

(4) The safety assessment shall be documented and revised by an independent expert where:

(a) modification of the radiation sources or its facilities are made :

(b) operational experience or information on accidents or errors indicates that the safety assessment shall be reviewed ; and

(c) techniques are modified in such a way that safety may be compromised.

Accident prevention.

61.—(1) The licenses shall incorporate in the safety procedures :

(a) defence-in-depth measures to cope with identified events, and evaluate the reliability of the safety systems ;

(b) the operational experience and lessons learned from accidents and errors into training, maintenance and quality assurance programmes ; and

(c) the licensee shall promptly inform the Authority of all reportable events.

Mitigation : Emergency plans.

62.—(1) Based on the events identified by the safety assessment, the licensee shall elaborate mitigation measures embodied in a set of emergency procedures and the relevant staff shall be trained in the mitigation measures, which shall be periodically rehearsed and the lessons learned from the rehearsals shall be used to review and update the emergency plans.

(2) The procedures shall identify the responsibilities of individuals and shall be concise, unambiguous and posted visibly in places where they could be needed.

(3) In cases where the beam control mechanism has failed to terminate the exposure at the end of the pre-set time, the procedure shall include the removal of the patient while avoiding exposure to the direct beam.

(4) A medical practitioner and medical physicist or competent staff specifically trained to remove the catheters and carry out the emergency procedures without delay shall be permanently present during High Dose Rate procedures and emergency equipment to retract the source to a fully shielded position shall also be available.

(5) Emergencies during source change shall be carried out only by maintenance staff trained and authorized for the tasks and if participation of the radiotherapy staff is necessary for any of these actions, the scope of this participation shall be restricted to the operation of the control panel and the responsibilities shall be clearly defined.

PART IX — SAFETY IN THE TRANSPORT OF
RADIOACTIVE MATERIALS

63.—(1) The licensee shall comply with the requirements of the Act, Nigerian Transportation of Radioactive Sources Regulations, 2006 and the International Atomic Energy Agency TS-R-1 for all activities involving transport of radioactive sources.

Safety
transportation
of
radioactive
materials.

(2) In the case of radiotherapy, the requirements apply to external beam radioactive sources and to Brachytherapy sources.

64. (1) Prior to each shipment of radioactive material to be dispatched, the licensee or the person responsible for the transport shall make the necessary arrangements with the source supplier, to receive the relevant information and the information shall include the following for each package or container:

Receipt of
radioactive
materials.

- (a) the nuclide, number and activity of sources ;
- (b) a description of the source construction and performance tests, including leakage tests ;
- (c) special form approval certificate where necessary ;
- (d) a description of the package ;
- (e) approval certificate for Type A or B packages, or statement of compliance with IAEA TS-R-1 for other packages ;
- (f) details of any special arrangements required, including multilateral approvals, where necessary ; and
- (g) a copy of the transport documents to be sent to the licensee by fax or e-mail before dispatch where necessary.

(2) The licensee shall not agree to the dispatch of the consignment by the supplier unless all the requirements of these regulations are satisfied and the supplier and licensee shall agree on the transport route and responsibility for each stage of the journey.

(3) Arrangements shall also be made for the following where necessary :

- (a) the need for special handling equipment for external beam sources, e.g., cranes, forklift trucks etc. during transfer from one mode of transport to another, or between vehicles ;

- (b) checking of radiation dose rates from the package or container ;
- (c) checking the correct transport labels as attached to the package or container, and replacing any that are damaged or illegible ;
- (d) ensuring that the package or container is securely attached to the vehicle and that the vehicle is correctly labeled ;
- (e) dealing with border controls ; and
- (f) security of the consignment during transport, particularly during delays or overnight stops.

Dispatch of radioactive materials.
Return of empty Packages.

65. The licensee shall return packages or containers to the source supplier after receipt of a consignment of radioactive material.

66. The licenses shall with regard to returning of empty packages :

- (a) carry out dose rate and contamination monitoring of both the inside and outside of the package or container to ensure that there is no residual radioactive material present and it can therefore be treated as an empty package or container ;
- (b) remove or cover all transport labels relating to the sources contained in the package or container when received ;
- (c) examine the package or container to ensure that it is in good condition, and then close it securely, referring to any procedures provided by the source supplier ;
- (d) attach a label to the outside of the package or container stating "UN 2908 RADIOACTIVE MATERIAL EXCEPTED PACKAGE — EMPTY PACKAGING" ;
- (e) complete a transport document ; and
- (f) contact the source supplier and agree the transport route and responsibility for each stage of the journey and inform the source supplier of the proposed date of dispatch.

Return of disused sources.

67.—(1) The licenses shall with regard to returning disused sources provide the following information to the consignee for each package or container :

- (a) the nuclide, number and activity of sources ;
- (b) a description of the source construction including leakage tests ;
- (c) special form approval certificate where necessary ;
- (d) a description of the packaging in which the source is to be transported ;
- (e) approval certificate for Type B package, or statement of compliance with International Atomic Energy Agency TS-R-1 for other packages where necessary.
- (f) details of any special arrangements required, including multilateral approvals, where necessary ; and
- (g) a copy of the transport documents to be sent to the consignee by fax or e-mail before dispatch where necessary.

(2) The licensee shall not dispatch the consignment unless they have received confirmation from the consignee that they are prepared to accept the consignment.

(3) The licensee and consignee shall agree on the transport route and responsibility for each stage of the journey and the Licensee will normally be responsible for dispatch until the consignment reaches the consignee's premises.

(4) Other arrangements are satisfactorily provided and they are agreed in advance by both parties and are also acceptable to the Authority and other regulatory authorities involved.

(5) The licenses shall in order to prepare the consignment for dispatch :

(a) load the sources into the package, verifying the details to be provided to the consignee e.g., serial numbers and comparable information to be entered on the transport document ;

(b) close it securely and then examine the package or container to ensure that it is in good condition, referring to any procedures provided by the source supplier ;

(c) carry out contamination monitoring of the outside of the package or container to ensure that there is no residual radioactive material present and it is therefore suitable for transport ;

(d) carry out dose rate monitoring of the package or container and attach appropriate transport labels ;

(e) refrain from using the transport labels relating to the sources contained in the package or container when received ;

(f) complete transport document ; and

(g) arrangements shall also be made for the following where necessary :

(i) specify the need for special handling equipment e.g., cranes, forklift trucks etc. during transfer from one mode of transport to another, or between vehicles ;

(ii) ensure that the package is securely attached to the vehicle and that the vehicle is correctly labeled ;

(iii) deal with border controls ; and

(iv) provide security of the consignment during transport, particularly during delays or overnight stops.

PART X — OFFENCES, PENALTIES AND APPEALS

68.—(1) Any person who contravenes any of the provisions of these regulations commits an offence.

Offences and penalties.

(2) Any one who commits an offence under these regulations shall be liable to the penalties as established in the enforcement policy issued by the Authority.

(3) The Authority shall impose penalties such as suspension, revocation of authorization, imposing administrative fine or closure of facility or any combination of these.

(4) Any person or body corporate who, being a holder of authorization under these regulations, commits an offence shall be liable to prosecution in the court of law and upon conviction be liable to pay fines not exceeding ₦1,000,000 for an individual and not exceeding ₦10,000,000 for a corporate body or be given a jail term not exceeding ten years or both.

69. Any person may appeal to the Board of the Authority against a decision made against him pursuant to these Regulations.

Appeal.

SCHEDULES

SCHEDULE 1

DOSE LIMITS FOR OCCUPATIONAL AND PUBLIC EXPOSURE

OCCUPATIONAL EXPOSURE

DOSE LIMITS

1.—(1) The occupational exposure of any worker shall be so controlled that the following limits be not exceeded :

- (a) an effective dose of 20 mSv per year averaged over five consecutive years
- (b) an effective dose of 50 mSv in any single year ;
- (c) an equivalent dose to the lens of the eye of 150 mSv in a year ; and
- (d) an equivalent dose to the extremities (hands and feet) or the skin of 500 mSv in a year.

(2) For apprentices of 16 to 18 years of age who are training for employment involving exposure to radiation and for students of age 16 to 18 who are required to use sources in the course of their studies, the occupational exposure shall be so controlled that the following limits be not exceeded :

- (a) an effective dose of 6 mSv in a year ;
- (b) an equivalent dose to the lens of the eye of 50 mSv in a year ; and
- (c) an equivalent dose to the extremities or the skin of 150 mSv in a year.

SPECIAL CIRCUMSTANCES

2. When, in special circumstances, a temporary change in the dose limitation requirements is approved pursuant to NiBIRR :

(a) the dose averaging period mentioned in para. (1.1)(a) may exceptionally be up to 10 consecutive years as specified by the Authority, and the effective dose for any worker shall not exceed 20 mSv per year averaged over this period and shall not exceed 50 mSv in any single year, and the circumstances shall be reviewed when the dose accumulated by any worker since the start of the extended averaging period reaches 100 mSv ; or

(b) the temporary change in the dose limitation shall be as specified by the Authority but shall not exceed 50 mSv in any year and the period of the temporary change shall not exceed 5 years.

PUBLIC EXPOSURE

DOSE LIMITS

3. The estimated average doses to the relevant critical groups of members of the public that are attributable to practices shall not exceed the following limits :

- (a) an effective dose of 1 mSv in a year ;
- (b) in special circumstances, an effective dose of up to 5 mSv in a single year provided that the average dose over five consecutive years does not exceed 1 mSv per year ;

- (c) an equivalent dose to the lens of the eye of 15 mSv in a year ; and
- (d) an equivalent dose to the skin of 50 mSv in a year.

DOSE LIMITATION FOR COMFORTERS AND VISITORS OF PATIENTS

4. The dose limits set out in this part shall not apply to comforters of patients, i.e., to individuals knowingly exposed while voluntarily helping (other than in their employment or occupation) in the care, support and comfort of patients undergoing medical diagnosis or treatment, or to visitors of such patients. However, the dose of any such comforter or visitor of patients shall be constrained so that it is unlikely that his or her dose will exceed 5 mSv during the period of a patient's diagnostic examination or treatment. The dose to children visiting patients who have ingested radioactive materials should be similarly constrained to less than 1 mSv.

MADE at Abuja this 5th day of October, 2006.

PROFESSOR SHAMSIDEEN BABATUNDE ELEGBA,
Director-General/Chief Executive Officer
Nigerian Nuclear Regulatory Authority

EXPLANATORY NOTE

(This note does not form part of the regulations but is intended to explain its purport)

1. This set of regulations is a practice-specific elaboration of the Nigeria Basic Ionizing Radiation Regulations, which is derived from, but not a substitute to, the International Basic Safety Standards for Protection against Ionizing Radiation Sources and published as International Atomic Energy Agency Safety Series No 115 in 1996.
2. The Regulations provide, amongst other things, for the protection of patients, workers and the public from the harmful effects of exposure to ionizing radiation.